



## **A novel and sensitive in situ instrument using incoherent broadband cavity-enhanced absorption spectroscopy for trace gas measurements**

A. A. Ruth (1,2), R. M. Varma (1,2,3), D. S. Venables (3,2), U. Heitmann (1), E. Schlosser (4), and S. Dixneuf (1)

(1) Department of Physics, University College Cork, Cork, Ireland (r.varma@ucc.ie), (2) Environmental Research Institute, University College Cork, Cork, Ireland, (3) Department of Chemistry, University College Cork, Cork, Ireland, (4) Forschungszentrum Jülich, ICG-2: Troposphäre, D-52425 Jülich, Germany

Incoherent broadband cavity-enhanced absorption spectroscopy (IBBCEAS) is a technique with high sensitivity to quantify trace gas concentrations based on their absorption features. The design and calibration of an incoherent broadband cavity-enhanced absorption spectroscopy (IBBCEAS) setup employing a 20 m long optical cavity is described for sensitive in situ measurements of light extinction between 630 nm and 690 nm. The setup was installed at the SAPHIR atmospheric simulation chamber during an intercomparison of instruments for nitrate (NO<sub>3</sub>) radical detection. The long cavity was stable for the entire duration of the two week campaign. A detection limit of ca. 2 pptv for NO<sub>3</sub> with a stable acquisition time of 5 s was achieved. In addition to monitoring NO<sub>3</sub>, nitrogen dioxide (NO<sub>2</sub>) concentrations were simultaneously retrieved and compared against concurrent measurements by a chemiluminescence detector. The spectral analysis of NO<sub>3</sub><sup>-</sup> and NO<sub>2</sub> from the SAPHIR chamber studies and comparison with a long path DOAS instrument are presented. Moreover, results from a field campaign at Roches Point, Ireland, will be discussed.