



Results of a European interlaboratory comparison on CO₂ sorption on activated carbon and coals

Yves Gensterblum (1), Andreas Busch (2), Bernhard Krooss (1), Guy de Weireld (3), Pierre Billemont (3), Patrick van Hemert (4), and Karl-Heinz Wolf (4)

(1) RWTH Aachen University, Institute of Geology and Geochemistry of Petroleum and Coal, Petrophysics, 52056, Germany (gensterblum@lekt.rwth-aachen.de), (2) Shell International Exploration and Production, Kessler Park 1, 2288 GS Rijswijk, The Netherlands, (3) Faculté Polytechnique de Mons, Service de Thermodynamique et Physique Mathématique, Mons, Belgium, (4) TU Delft, Department of Geotechnology, Delft, Netherlands, (5) INERIS / LCA-ENSIACET, Laboratory of agro-industrial chemistry, Paris, France

For the assessment of CO₂ storage in coal seams or enhanced coalbed methane production (ECBM), the sorption properties of natural coals are important parameters. Since more and more laboratories worldwide are concerned with measurements of gas sorption on coal it is indispensable to establish quality standards for such experiments. The first two interlaboratory studies on CO₂ sorption on coal (Goodman et al. 2004, 2007) revealed a poor agreement of sorption isotherms among the participating laboratories, particularly in the high-pressure range. During the MOVECBM (<http://www.movecbm.eu/>) project funded by the European Commission (6th framework), an interlaboratory comparison of CO₂ sorption on selected coals and activated carbon was initiated. Measurements were performed on dry samples at 45° C using the manometric and the gravimetric method, up to a final pressure of 15 MPa.

The first set of high-pressure sorption measurements was performed on a Filtrasorb 400 activated carbon sample in order to minimise heterogeneity effects and to optimize the experimental procedures for the individual (manometric or gravimetric) methods (Gensterblum et al. 2009).

Since comparability for the activated carbon was excellent, the measurements were continued using natural coals of various rank (anthracite, bituminous coal and lignite) to study the influence of heterogeneities and varying starting conditions on the CO₂ sorption properties (Gensterblum et al. 2010).

Compared to the poor reproducibility observed in previous interlaboratory studies (Goodman et al., 2004, 2007) this European study showed excellent agreement (<5 % deviation) among the participating laboratories with good repeatability.

The sorption data and technical information on the different experimental setups have been used to investigate errors and potential pitfalls in the assessment of high-pressure CO₂ sorption isotherms.

References

Gensterblum Y., P. van Hemert, P. Billemont, A. Busch, B.M. Krooss, G. de Weireld, D. Prinz , K.-H.A.A. Wolf, "European inter-laboratory comparison of high pressure CO₂ sorption isotherms. II: natural coals" IJCG, 2010, 84, 115-124

Gensterblum Y., P. van Hemert, P. Billemont, A. Busch, D. Charrière, D. Li, B.M. Krooss, G. de Weireld, D. Prinz , K.-H.A.A. Wolf, "European inter-laboratory comparison of high pressure CO₂ sorption isotherms. I: Activated carbon" Carbon 47 (2009) 2958 –2969

Goodman, A.L., Busch, A., Duffy, G., Fitzgerald, J.E., Gasem, K.A.M., Gensterblum, Y., Krooss, B.M., Levy, J., Ozdemir, E., Pan, Z., Robinson, Jr., R.L., Schroeder, K., Sudibandriyo, M., White, C. (2004). An Inter-laboratory Comparison of CO₂ Isotherms Measured on Argonne Premium Coal Samples. Energy and Fuels 18, 1175-1182.

Goodman, A.L., Busch, A., Day, S., Duffy, G.J., Fitzgerald, J.E., Gasem, K.A.M., Gensterblum, Y., Hartman, C., Krooss, B.M., Pan, Z., Pratt, T., Robinson, Jr., R.L., Romanov, V., Sakurovs, R., Schroeder, K., Sudibandriyo, M., White, C.M. (2007) "Inter-laboratory Comparison II: CO₂ Isotherms Measured on Moisture-Equilibrated Argonne Premium Coals at 55oC and 15 MPa", International Journal of Coal Geology 72, 153-164.